

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024210.D
 Acq On : 14 Sep 2021 16:17
 Operator : JC/MD
 Sample : M3729-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-EB-20210909

Quant Time: Sep 15 01:57:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135775	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231615	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	99176	51.224	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.440%	
35) Dibromofluoromethane	5.397	113	77576	49.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.880%	
50) Toluene-d8	8.653	98	283719	50.139	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.280%	
62) 4-Bromofluorobenzene	11.085	95	98441	44.702	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.400%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.301	50	856	0.628	ug/l	91
11) Tert butyl alcohol	2.971	59	21075	33.275	ug/l	97
16) Acetone	2.386	43	24637	29.843	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024210.D
Acq On : 14 Sep 2021 16:17
Operator : JC/MD
Sample : M3729-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-EB-20210909

Quant Time: Sep 15 01:57:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

