

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024212.D
 Acq On : 14 Sep 2021 17:04
 Operator : JC/MD
 Sample : M3729-07DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-760-762DL

Quant Time: Sep 15 01:57:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235447	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	224600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98412	50.640	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.280%
35) Dibromofluoromethane	5.397	113	75052	47.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	8.653	98	287611	50.000	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.000%
62) 4-Bromofluorobenzene	11.085	95	109861	49.076	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.160%
Target Compounds						
						Qvalue
16) Acetone	2.398	43	2845	3.433	ug/l #	83
44) Trichloroethene	7.135	130	8342	4.782	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024212.D
Acq On : 14 Sep 2021 17:04
Operator : JC/MD
Sample : M3729-07DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-760-762DL

Quant Time: Sep 15 01:57:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

