

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024214.D
 Acq On : 14 Sep 2021 17:52
 Operator : JC/MD
 Sample : M3729-09DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-DUP-20210909DL

Quant Time: Sep 15 01:57:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230837	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	218914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95192	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	99693	51.785	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.580%
35) Dibromofluoromethane	5.397	113	77750	50.223	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.440%
50) Toluene-d8	8.653	98	283149	50.207	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.420%
62) 4-Bromofluorobenzene	11.085	95	105061	47.869	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.740%
Target Compounds						
16) Acetone	2.386	43	2928	3.567	ug/l #	81
44) Trichloroethene	7.135	130	11234	6.569	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024214.D
Acq On : 14 Sep 2021 17:52
Operator : JC/MD
Sample : M3729-09DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-DUP-20210909DL

Quant Time: Sep 15 01:57:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

