

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024220.D
 Acq On : 14 Sep 2021 20:13
 Operator : JC/MD
 Sample : M3687-37
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-210909

Quant Time: Sep 15 01:59:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95098	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102146	52.594	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	105.180%	
35) Dibromofluoromethane	5.397	113	78472	50.204	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.400%	
50) Toluene-d8	8.653	98	285216	50.089	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.180%	
62) 4-Bromofluorobenzene	11.085	95	104886	47.332	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.660%	
Target Compounds						
16) Acetone	2.392	43	1134	1.369	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024220.D
Acq On : 14 Sep 2021 20:13
Operator : JC/MD
Sample : M3687-37
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-210909

Quant Time: Sep 15 01:59:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

