

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024228.D
 Acq On : 15 Sep 2021 00:08
 Operator : JC/MD
 Sample : PB139075ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139075ZHE#01

Quant Time: Sep 15 02:00:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	133474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	220407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99677	52.371	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.740%			
35) Dibromofluoromethane	5.397	113	77528	50.671	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.340%			
50) Toluene-d8	8.653	98	285655	51.249	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.500%			
62) 4-Bromofluorobenzene	11.085	95	102517	47.261	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.520%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	46513	57.314	ug/l #	85
20) Methylene Chloride	2.794	84	19070	12.237	ug/l	90
25) 2-Butanone	4.574	43	8387	7.017	ug/l	95
43) Isopropyl Acetate	6.348	43	18865	4.849	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024228.D
Acq On : 15 Sep 2021 00:08
Operator : JC/MD
Sample : PB139075ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139075ZHE#01

Quant Time: Sep 15 02:00:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

