

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024229.D
 Acq On : 15 Sep 2021 00:32
 Operator : JC/MD
 Sample : PB139075ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139075ZHE#02

Quant Time: Sep 15 02:01:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	136174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	230855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96704	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102143	52.602	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.200%			
35) Dibromofluoromethane	5.397	113	78276	49.644	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.280%			
50) Toluene-d8	8.653	98	294106	51.202	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.400%			
62) 4-Bromofluorobenzene	11.085	95	109334	48.911	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.820%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	46728	56.437	ug/l	90
20) Methylene Chloride	2.794	84	19591	12.322	ug/l	91
25) 2-Butanone	4.580	43	8034	6.588	ug/l	92
43) Isopropyl Acetate	6.354	43	19305	4.815	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024229.D
Acq On : 15 Sep 2021 00:32
Operator : JC/MD
Sample : PB139075ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139075ZHE#02

Quant Time: Sep 15 02:01:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

