

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024230.D
 Acq On : 15 Sep 2021 00:56
 Operator : JC/MD
 Sample : PB139075ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139075ZHE#03

Quant Time: Sep 15 02:01:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218466	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	96290	52.947	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.900%
35) Dibromofluoromethane	5.397	113	72398	49.414	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.820%
50) Toluene-d8	8.653	98	277502	51.992	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.980%
62) 4-Bromofluorobenzene	11.085	95	94689	45.586	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.180%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	28956	37.341	ug/l	88
20) Methylene Chloride	2.794	84	9284	6.235	ug/l	90
43) Isopropyl Acetate	6.355	43	21424	5.751	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024230.D
Acq On : 15 Sep 2021 00:56
Operator : JC/MD
Sample : PB139075ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139075ZHE#03

Quant Time: Sep 15 02:01:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

