

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024232.D
 Acq On : 15 Sep 2021 01:43
 Operator : JC/MD
 Sample : PB139076TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139076TB

Quant Time: Sep 15 07:30:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	224653	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	217467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96017	51.538	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.080%
35) Dibromofluoromethane	5.397	113	75406	50.050	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.100%
50) Toluene-d8	8.653	98	284245	51.789	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.580%
62) 4-Bromofluorobenzene	11.085	95	97728	45.754	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.500%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	29828	37.548	ug/l #	87
20) Methylene Chloride	2.794	84	9597	6.291	ug/l	93
43) Isopropyl Acetate	6.355	43	22434	5.856	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024232.D
Acq On : 15 Sep 2021 01:43
Operator : JC/MD
Sample : PB139076TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139076TB

Quant Time: Sep 15 07:30:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

