

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024236.D
 Acq On : 15 Sep 2021 03:17
 Operator : JC/MD
 Sample : M3724-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31191

Quant Time: Sep 15 07:31:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	125761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	216438	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	206622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	95123	53.043	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.080%
35) Dibromofluoromethane	5.397	113	73706	50.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.560%
50) Toluene-d8	8.653	98	273002	51.628	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.260%
62) 4-Bromofluorobenzene	11.085	95	92556	44.977	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.960%
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	30906	40.418	ug/l	89
20) Methylene Chloride	2.800	84	10677	7.271	ug/l #	94
43) Isopropyl Acetate	6.348	43	21046	5.702	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	2763	1.024	ug/l	99
95) Naphthalene	13.780	128	6894	2.433	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024236.D
Acq On : 15 Sep 2021 03:17
Operator : JC/MD
Sample : M3724-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31191

Quant Time: Sep 15 07:31:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

