

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024242.D
 Acq On : 15 Sep 2021 05:38
 Operator : JC/MD
 Sample : M3724-16 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33132

Quant Time: Sep 15 07:33:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	132454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229115	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91345	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	100484	53.201	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.400%
35) Dibromofluoromethane	5.397	113	76732	49.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.880%
50) Toluene-d8	8.652	98	279781	49.983	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.960%
62) 4-Bromofluorobenzene	11.085	95	100811	46.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024242.D
Acq On : 15 Sep 2021 05:38
Operator : JC/MD
Sample : M3724-16 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33132

Quant Time: Sep 15 07:33:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

