

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024244.D
 Acq On : 15 Sep 2021 06:25
 Operator : JC/MD
 Sample : M3724-18 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33134

Quant Time: Sep 15 07:34:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231233	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95815	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101847	54.211	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.420%
35) Dibromofluoromethane	5.397	113	77354	49.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.760%
50) Toluene-d8	8.653	98	286742	50.757	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.520%
62) 4-Bromofluorobenzene	11.085	95	104077	47.340	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024244.D
Acq On : 15 Sep 2021 06:25
Operator : JC/MD
Sample : M3724-18 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33134

Quant Time: Sep 15 07:34:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

