

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024246.D
 Acq On : 15 Sep 2021 07:12
 Operator : JC/MD
 Sample : M3724-20 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33136

Quant Time: Sep 15 07:57:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229079	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	217655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93891	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100801	53.202	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.400%
35) Dibromofluoromethane	5.397	113	77512	50.454	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.900%
50) Toluene-d8	8.653	98	281786	50.349	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.700%
62) 4-Bromofluorobenzene	11.085	95	101879	46.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024246.D
Acq On : 15 Sep 2021 07:12
Operator : JC/MD
Sample : M3724-20 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33136

Quant Time: Sep 15 07:57:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

