

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
 Data File : VX024238.D
 Acq On : 15 Sep 2021 04:04
 Operator : JC/MD
 Sample : M3724-12 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33128

Quant Time: Sep 15 07:32:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229511	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218363	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	100523	52.786	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.580%	
35) Dibromofluoromethane	5.397	113	75800	49.246	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.500%	
50) Toluene-d8	8.653	98	282608	50.401	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.800%	
62) 4-Bromofluorobenzene	11.085	95	103331	47.353	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.700%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091421\
Data File : VX024238.D
Acq On : 15 Sep 2021 04:04
Operator : JC/MD
Sample : M3724-12 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33128

Quant Time: Sep 15 07:32:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04




Instrument :
MSVOA_X
ClientSampleId :
33128


Tgt Ion:168 Resp: 133549
Ion Ratio Lower Upper
168 100
99 60.1 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 52.786 ug/l
RT: 5.971 min Scan# 801
Delta R.T. -0.000 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04

Tgt Ion: 65 Resp: 100523
Ion Ratio Lower Upper
65 100
67 49.3 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04




Instrument :
MSVOA_X
ClientSampled :
33128


Tgt Ion:114 Resp: 229511
Ion Ratio Lower Upper
114 100
63 20.8 0.0 40.8
88 16.4 0.0 32.8



#35
Dibromofluoromethane
Concen: 49.246 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04

Tgt Ion:113 Resp: 75800
Ion Ratio Lower Upper
113 100
111 104.3 81.9 122.9
192 19.9 16.7 25.1









#50
Toluene-d8
Concen: 50.401 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04




Instrument :
MSVOA_X
ClientSampled :
33128


Tgt Ion: 98 Resp: 282608
Ion Ratio Lower Upper
98 100
100 63.0 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 47.353 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04

Tgt Ion: 95 Resp: 103331
Ion Ratio Lower Upper
95 100
174 81.5 0.0 154.6
176 77.6 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04

Instrument :
MSVOA_X
ClientSampled :
33128

Tgt Ion:117 Resp: 218363
Ion Ratio Lower Upper
117 100
82 58.4 47.4 71.2
119 30.9 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024238.D
Acq: 15 Sep 2021 04:04

Tgt Ion:152 Resp: 94925
Ion Ratio Lower Upper
152 100
115 60.7 41.1 123.3
150 158.9 0.0 349.0

