

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
 Data File : VX031322.D
 Acq On : 14 Sep 2022 21:48
 Operator : JC/MD
 Sample : N4497-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 15 06:06:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	61040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	111860	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66333	44.274	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.540%
35) Dibromofluoromethane	5.385	113	47679	39.653	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	79.300%
50) Toluene-d8	8.653	98	187103	41.482	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	82.960%#
62) 4-Bromofluorobenzene	11.079	95	62153	37.630	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.260%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
Data File : VX031322.D
Acq On : 14 Sep 2022 21:48
Operator : JC/MD
Sample : N4497-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 15 06:06:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

