

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
 Data File : VX031326.D
 Acq On : 14 Sep 2022 23:22
 Operator : JC/MD
 Sample : N4446-07RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-WATER-LIQUIDRE

Quant Time: Sep 15 06:07:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	60118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	108340	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39422	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69420	47.045	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.080%		
35) Dibromofluoromethane	5.385	113	47387	40.690	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	81.380%		
50) Toluene-d8	8.653	98	141880	32.478	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.960%#		
62) 4-Bromofluorobenzene	11.079	95	45221	28.268	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	56.540%#		
Target Compounds					Qvalue	
16) Acetone	2.398	43	5540	8.794	ug/l	90
25) 2-Butanone	4.580	43	4149	3.993	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
Data File : VX031326.D
Acq On : 14 Sep 2022 23:22
Operator : JC/MD
Sample : N4446-07RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-WATER-LIQUIDRE

Quant Time: Sep 15 06:07:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

