

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
 Data File : VX031327.D
 Acq On : 14 Sep 2022 23:45
 Operator : JC/MD
 Sample : N4446-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-YRE

Quant Time: Sep 15 06:08:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	60887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	109765	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69422	46.452	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.900%
35) Dibromofluoromethane	5.391	113	46580	39.478	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	78.960%
50) Toluene-d8	8.652	98	142475	32.191	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	64.380%#
62) 4-Bromofluorobenzene	11.079	95	46180	28.493	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	56.980%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
Data File : VX031327.D
Acq On : 14 Sep 2022 23:45
Operator : JC/MD
Sample : N4446-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-YRE

Quant Time: Sep 15 06:08:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

