

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
 Data File : VX031329.D
 Acq On : 15 Sep 2022 00:32
 Operator : JC/MD
 Sample : N4578-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHERRY-HILL-COMP-5

Quant Time: Sep 15 06:08:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	59433	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	108982	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65023	44.573	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.140%		
35) Dibromofluoromethane	5.391	113	47011	40.130	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	80.260%		
50) Toluene-d8	8.653	98	175155	39.859	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	79.720%#		
62) 4-Bromofluorobenzene	11.079	95	58298	36.228	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	72.460%#		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	18301	29.386	ug/l	97
20) Methylene Chloride	2.788	84	5809	3.738	ug/l	96
37) Ethyl Acetate	4.727	43	12503	5.934	ug/l	97
43) Isopropyl Acetate	6.342	43	53237	14.801	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
Data File : VX031329.D
Acq On : 15 Sep 2022 00:32
Operator : JC/MD
Sample : N4578-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHERRY-HILL-COMP-5

Quant Time: Sep 15 06:08:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

