

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
 Data File : VX031345.D
 Acq On : 15 Sep 2022 07:56
 Operator : JC/MD
 Sample : N4579-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHERRY-HILL-COMP-14

Quant Time: Sep 15 11:13:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	54272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	100043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38411	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57740	43.344	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.680%
35) Dibromofluoromethane	5.391	113	42493	39.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	79.020%
50) Toluene-d8	8.653	98	167936	41.631	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	83.260%#
62) 4-Bromofluorobenzene	11.079	95	56864	38.495	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	76.980%#
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9357	16.454	ug/l	97
20) Methylene Chloride	2.794	84	3269	1.342	ug/l	90
37) Ethyl Acetate	4.727	43	11769	6.084	ug/l #	88
43) Isopropyl Acetate	6.342	43	52781	15.986	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
Data File : VX031345.D
Acq On : 15 Sep 2022 07:56
Operator : JC/MD
Sample : N4579-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHERRY-HILL-COMP-14

Quant Time: Sep 15 11:13:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

