

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
 Data File : VX031348.D
 Acq On : 15 Sep 2022 09:06
 Operator : JC/MD
 Sample : N4539-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MBD-WC-S-MP-1-3-4

Quant Time: Sep 15 11:13:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	56091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	100777	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68440	49.710	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.420%
35) Dibromofluoromethane	5.385	113	49382	45.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.180%
50) Toluene-d8	8.653	98	190359	46.846	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.700%
62) 4-Bromofluorobenzene	11.079	95	63144	42.434	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.860%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	6593	11.217	ug/l	92
37) Ethyl Acetate	4.727	43	12099	6.209	ug/l #	94
43) Isopropyl Acetate	6.342	43	50306	15.125	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
Data File : VX031348.D
Acq On : 15 Sep 2022 09:06
Operator : JC/MD
Sample : N4539-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MBD-WC-S-MP-1-3-4

Quant Time: Sep 15 11:13:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

