



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091422\
Data File : VX031311.D
Acq On : 14 Sep 2022 16:20
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Title : SW846 8260
Last Update : Thu Sep 15 01:50:07 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.3	5648	PASS
75	95	30	60	58.4	13585	PASS
95	95	100	100	100.0	23281	PASS
96	95	5	9	7.0	1626	PASS
173	174	0.00	2	0.8	123	PASS
174	95	50	100	65.0	15141	PASS
175	174	5	9	8.0	1206	PASS
176	174	95	101	98.1	14847	PASS
177	176	5	9	6.8	1012	PASS