

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091518\
 Data File : VX004605.D
 Acq On : 16 Sep 2018 01:19
 Operator : JC/MD
 Sample : J4913-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPSI-TT-MW30211-20180912

Quant Time: Sep 17 17:50:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	255152	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	349266	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	209701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165158	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
35) Dibromofluoromethane	5.50	113	151243	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	8.71	98	494696	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
62) 4-Bromofluorobenzene	11.14	95	174813	45.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.38%	
Target Compounds						
24) 1,1-Dichloroethane	3.70	63	9793	1.76	ug/l	95
44) Trichloroethene	7.21	130	14323	4.38	ug/l	90
64) Tetrachloroethene	9.34	164	3198	0.77	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091518\
Data File : VX004605.D
Acq On : 16 Sep 2018 01:19
Operator : JC/MD
Sample : J4913-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPSI-TT-MW302I1-20180912

Quant Time: Sep 17 17:50:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

