

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091518\
 Data File : VX004611.D
 Acq On : 16 Sep 2018 03:59
 Operator : JC/MD
 Sample : J4913-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW30212-20180912

Quant Time: Sep 17 09:06:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	272341	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	337396	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	317422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199632	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165026	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
35) Dibromofluoromethane	5.49	113	152281	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	8.72	98	479975	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
62) 4-Bromofluorobenzene	11.14	95	165030	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	1838	0.61	ug/l	95
16) Acetone	2.44	43	2681	1.56	ug/l	91
44) Trichloroethene	7.21	130	20127	6.37	ug/l	99
64) Tetrachloroethene	9.34	164	1361	0.34	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091518\
Data File : VX004611.D
Acq On : 16 Sep 2018 03:59
Operator : JC/MD
Sample : J4913-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW30212-20180912

Quant Time: Sep 17 09:06:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

