

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091518\
 Data File : VX004613.D
 Acq On : 16 Sep 2018 04:52
 Operator : JC/MD
 Sample : J4913-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW302D-20180913

Quant Time: Sep 17 09:03:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	259198	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	363404	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	334098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	228281	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	173632	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
35) Dibromofluoromethane	5.51	113	161575	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
50) Toluene-d8	8.72	98	498239	44.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.36%	
62) 4-Bromofluorobenzene	11.14	95	174669	43.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3542	2.17	ug/l	90
44) Trichloroethene	7.22	130	7621	2.24	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091518\
Data File : VX004613.D
Acq On : 16 Sep 2018 04:52
Operator : JC/MD
Sample : J4913-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW302D-20180913

Quant Time: Sep 17 09:03:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

