

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091518\
 Data File : VX004614.D
 Acq On : 16 Sep 2018 05:19
 Operator : JC/MD
 Sample : J4913-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW3071-20180913

Quant Time: Sep 17 09:02:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	250438	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	335675	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	322336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208304	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	166208	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
35) Dibromofluoromethane	5.51	113	151795	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
50) Toluene-d8	8.72	98	471014	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
62) 4-Bromofluorobenzene	11.14	95	168615	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	3222	2.04	ug/l	99
44) Trichloroethene	7.21	130	3145	1.00	ug/l	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091518\
Data File : VX004614.D
Acq On : 16 Sep 2018 05:19
Operator : JC/MD
Sample : J4913-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW307I-20180913

Quant Time: Sep 17 09:02:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

