

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091518\
 Data File : VX004597.D
 Acq On : 15 Sep 2018 20:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 17 08:44:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.469	0.447	4.7	85	0.00
3 P	Chloromethane	0.645	0.571	11.5	87	0.00
4 C	Vinyl Chloride	0.662	0.628	5.1#	88	0.00
5 T	Bromomethane	0.316	0.223	29.4#	75	0.00
6 T	Chloroethane	0.428	0.420	1.9	94	0.00
7 T	Trichlorofluoromethane	0.877	0.820	6.5	91	0.00
8 T	Diethyl Ether	0.393	0.399	-1.5	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.534	4.1	94	0.00
10 T	Methyl Iodide	0.612	0.397	35.1#	56	0.00
11 T	Tert butyl alcohol	0.154	0.141	8.4	89	0.00
12 CM	1,1-Dichloroethene	0.522	0.512	1.9#	94	0.00
13 T	Acrolein	0.061	0.070	-14.8	122	0.00
14 T	Allyl chloride	1.074	0.978	8.9	86	0.00
15 T	Acrylonitrile	0.370	0.350	5.4	94	0.00
16 T	Acetone	0.315	0.262	16.8	87	0.00
17 T	Carbon Disulfide	1.510	1.467	2.8	90	0.00
18 T	Methyl Acetate	0.883	0.859	2.7	97	0.00
19 T	Methyl tert-butyl Ether	1.833	1.880	-2.6	100	0.00
20 T	Methylene Chloride	0.645	0.604	6.4	95	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.551	4.5	93	0.00
22 T	Diisopropyl ether	1.934	1.929	0.3	96	0.00
23 T	Vinyl Acetate	1.565	1.583	-1.2	91	0.00
24 P	1,1-Dichloroethane	1.092	1.076	1.5	94	0.00
25 T	2-Butanone	0.501	0.458	8.6	92	0.00
26 T	2,2-Dichloropropane	0.750	0.634	15.5	82	0.00
27 T	cis-1,2-Dichloroethene	0.644	0.655	-1.7	98	0.00
28 T	Bromochloromethane	0.500	0.521	-4.2	101	0.00
29 T	Tetrahydrofuran	0.302	0.299	1.0	91	-0.01
30 C	Chloroform	1.066	1.038	2.6#	92	0.00
31 T	Cyclohexane	0.933	0.971	-4.1	91	0.00
32 T	1,1,1-Trichloroethane	0.859	0.854	0.6	96	0.00
33 S	1,2-Dichloroethane-d4	0.645	0.646	-0.2	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.458	0.441	3.7	92	0.00
36 T	1,1-Dichloropropene	0.607	0.623	-2.6	97	0.00
37 T	Ethyl Acetate	0.690	0.666	3.5	91	0.00
38 T	Carbon Tetrachloride	0.593	0.597	-0.7	94	0.00
39 T	Methylcyclohexane	0.633	0.658	-3.9	94	0.00
40 TM	Benzene	1.801	1.777	1.3	92	0.00
41 T	Methacrylonitrile	0.380	0.376	1.1	95	0.00
42 TM	1,2-Dichloroethane	0.597	0.569	4.7	96	-0.01
43 T	Isopropyl Acetate	0.917	0.935	-2.0	98	0.00
44 TM	Trichloroethene	0.468	0.456	2.6	94	0.00
45 C	1,2-Dichloropropane	0.433	0.434	-0.2#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091518\
 Data File : VX004597.D
 Acq On : 15 Sep 2018 20:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 17 08:44:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.274	0.272	0.7	95	0.00
47 T	Bromodichloromethane	0.518	0.518	0.0	95	0.00
48 T	Methyl methacrylate	0.454	0.469	-3.3	94	0.00
49 T	1,4-Dioxane	0.011	0.011	0.0	90	0.00
50 S	Toluene-d8	1.534	1.498	2.3	95	0.00
51 T	4-Methyl-2-Pentanone	0.647	0.598	7.6	93	0.00
52 CM	Toluene	1.061	1.046	1.4#	92	0.00
53 T	t-1,3-Dichloropropene	0.605	0.581	4.0	84	0.00
54 T	cis-1,3-Dichloropropene	0.617	0.639	-3.6	95	0.00
55 T	1,1,2-Trichloroethane	0.457	0.420	8.1	86	0.00
56 T	Ethyl methacrylate	0.646	0.647	-0.2	87	0.00
57 T	1,3-Dichloropropane	0.742	0.702	5.4	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.323	-8.8	95	0.00
59 T	2-Hexanone	0.496	0.454	8.5	90	0.00
60 T	Dibromochloromethane	0.430	0.428	0.5	96	0.00
61 T	1,2-Dibromoethane	0.436	0.434	0.5	96	0.00
62 S	4-Bromofluorobenzene	0.548	0.553	-0.9	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.624	0.557	10.7	82	0.00
65 PM	Chlorobenzene	1.242	1.243	-0.1	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.428	0.440	-2.8	97	0.00
67 C	Ethyl Benzene	2.019	2.103	-4.2#	96	0.00
68 T	m/p-Xylenes	0.778	0.823	-5.8	95	0.00
69 T	o-Xylene	0.748	0.801	-7.1	98	0.00
70 T	Styrene	1.239	1.339	-8.1	96	0.00
71 P	Bromoform	0.352	0.362	-2.8	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.293	3.568	-8.4	97	0.00
74 T	N-amyl acetate	1.386	1.504	-8.5	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.062	1.084	-2.1	100	0.00
76 T	1,2,3-Trichloropropane	1.006	1.016	-1.0	95	0.00
77 T	Bromobenzene	0.910	0.945	-3.8	98	0.00
78 T	n-propylbenzene	3.786	4.070	-7.5	95	0.00
79 T	2-Chlorotoluene	2.296	2.396	-4.4	96	0.00
80 T	1,3,5-Trimethylbenzene	2.768	2.989	-8.0	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.316	0.315	0.3	90	0.00
82 T	4-Chlorotoluene	2.680	2.797	-4.4	94	0.00
83 T	tert-Butylbenzene	2.635	2.920	-10.8	99	0.00
84 T	1,2,4-Trimethylbenzene	2.826	3.062	-8.4	95	0.00
85 T	sec-Butylbenzene	3.358	3.634	-8.2	95	0.00
86 T	p-Isopropyltoluene	2.966	3.240	-9.2	95	0.00
87 T	1,3-Dichlorobenzene	1.698	1.732	-2.0	96	0.00
88 T	1,4-Dichlorobenzene	1.755	1.751	0.2	96	0.00
89 T	n-Butylbenzene	2.703	2.779	-2.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091518\
Data File : VX004597.D
Acq On : 15 Sep 2018 20:53
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 17 08:44:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.475	0.504	-6.1	99	0.00
91 T	1,2-Dichlorobenzene	1.719	1.768	-2.9	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.241	-6.2	94	0.00
93 T	1,2,4-Trichlorobenzene	1.217	1.312	-7.8	97	0.00
94 T	Hexachlorobutadiene	0.636	0.644	-1.3	97	0.00
95 T	Naphthalene	3.413	4.372	-28.1#	109	0.00
96 T	1,2,3-Trichlorobenzene	1.250	1.343	-7.4	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6