

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091518\
 Data File : VX004616.D
 Acq On : 16 Sep 2018 06:12
 Operator : JC/MD
 Sample : J4913-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW303S-20180913

Quant Time: Sep 19 16:20:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	264164	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	365357	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	342048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	236631	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174256	51.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.32%	
35) Dibromofluoromethane	5.49	113	161593	48.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.66%	
50) Toluene-d8	8.72	98	500834	44.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.34%	
62) 4-Bromofluorobenzene	11.14	95	178294	44.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.10%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091518\
Data File : VX004616.D
Acq On : 16 Sep 2018 06:12
Operator : JC/MD
Sample : J4913-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW303S-20180913

Quant Time: Sep 19 16:20:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX004616.D

Acq: 16 Sep 2018 06:12

Instrument :

MSVOA_X

ClientSampleId :


BPS1-TT-MW303S-20180913

Tot Ion:168 Resp: 264164

Ion Ratio Lower Upper




168 100

99 49.7 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 51.16 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX004616.D


Acq: 16 Sep 2018 06:12

Tgt Ion: 65 Resp: 174256

Ion Ratio Lower Upper



65 100


67 55.1 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX004616.D

Acq: 16 Sep 2018 06:12

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW303S-20180913


Tot Ion:114 Resp: 365357

Ion Ratio Lower Upper

114 100

63 18.2 0.0 39.0




88 14.9 0.0 30.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.33 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.01 min

Lab File: VX004616.D

Acq: 16 Sep 2018 06:12


Tgt Ion:113 Resp: 161593

Ion Ratio Lower Upper

113 100

111 103.8 82.4 123.6



192 23.8 19.4 29.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 44.67 ug/l
RT: 8.72 min Scan# 1241
Delta R.T. -0.00 min
Lab File: VX004616.D
Acq: 16 Sep 2018 06:12




Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW303S-20180913


Tot Ion: 98 Resp: 500834
Ion Ratio Lower Upper
98 100
100 64.3 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 44.55 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. -0.00 min
Lab File: VX004616.D
Acq: 16 Sep 2018 06:12

Tgt Ion: 95 Resp: 178294
Ion Ratio Lower Upper
95 100
174 94.9 0.0 185.2
176 91.6 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX004616.D
Acq: 16 Sep 2018 06:12

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW303S-20180913

Tot Ion:117 Resp: 342048
Ion Ratio Lower Upper
117 100
82 51.7 47.8 71.6
119 32.0 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX004616.D
Acq: 16 Sep 2018 06:12

Tgt Ion:152 Resp: 236631
Ion Ratio Lower Upper
152 100
115 54.9 39.0 117.0
150 148.5 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

