

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091520\
 Data File : VX018405.D
 Acq On : 15 Sep 2020 14:38
 Operator : JC/SP
 Sample : L3980-01 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31239

Quant Time: Sep 16 03:59:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	460834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	733370	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	699393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	349300	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	323980	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
35) Dibromofluoromethane	5.46	113	252649	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
50) Toluene-d8	8.70	98	891323	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
62) 4-Bromofluorobenzene	11.12	95	331868	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
Target Compounds						
16) Acetone	2.42	43	74455	29.867	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091520\
Data File : VX018405.D
Acq On : 15 Sep 2020 14:38
Operator : JC/SP
Sample : L3980-01 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31239

Quant Time: Sep 16 03:59:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

