

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091520\
 Data File : VX018413.D
 Acq On : 15 Sep 2020 17:39
 Operator : JC/SP
 Sample : PB131607ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131607ZHE#04

Quant Time: Sep 16 04:17:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	427267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	706889	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	705393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	344144	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	321020	54.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.24%	
35) Dibromofluoromethane	5.47	113	248340	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
50) Toluene-d8	8.70	98	897532	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.12	95	336568	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
Target Compounds						
16) Acetone	2.42	43	21137	9.145	ug/l	92
18) Methyl Acetate	2.75	43	7874	1.412	ug/l	94
20) Methylene Chloride	2.84	84	17287	2.550	ug/l	90
43) Isopropyl Acetate	6.42	43	119139	9.741	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091520\
Data File : VX018413.D
Acq On : 15 Sep 2020 17:39
Operator : JC/SP
Sample : PB131607ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131607ZHE#04

Quant Time: Sep 16 04:17:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

