

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091520\
 Data File : VX018415.D
 Acq On : 15 Sep 2020 18:24
 Operator : JC/SP
 Sample : PB131607ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131607ZHE#06

Quant Time: Sep 16 04:22:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	426533	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	698517	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	705212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	342828	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	312573	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
35) Dibromofluoromethane	5.46	113	245003	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
50) Toluene-d8	8.70	98	893128	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	11.12	95	331447	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
Target Compounds						
16) Acetone	2.42	43	22437	9.724	ug/l	95
18) Methyl Acetate	2.76	43	7290	1.309	ug/l #	86
20) Methylene Chloride	2.84	84	15292	2.123	ug/l	96
43) Isopropyl Acetate	6.42	43	120562	9.975	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091520\
Data File : VX018415.D
Acq On : 15 Sep 2020 18:24
Operator : JC/SP
Sample : PB131607ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131607ZHE#06

Quant Time: Sep 16 04:22:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

