

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091520\
 Data File : VX018421.D
 Acq On : 15 Sep 2020 20:41
 Operator : JC/SP
 Sample : L4005-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 32-34

Quant Time: Sep 16 04:31:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	422054	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	699533	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	688593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	335132	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	315426	53.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.66%	
35) Dibromofluoromethane	5.47	113	245243	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
50) Toluene-d8	8.70	98	848214	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
62) 4-Bromofluorobenzene	11.12	95	319596	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	
Target Compounds						
16) Acetone	2.42	43	186846	81.839	ug/l	92
18) Methyl Acetate	2.74	43	5978	1.085	ug/l #	85
20) Methylene Chloride	2.84	84	13527	1.771	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091520\
Data File : VX018421.D
Acq On : 15 Sep 2020 20:41
Operator : JC/SP
Sample : L4005-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
32-34

Quant Time: Sep 16 04:31:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

