

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091520\
 Data File : VX018423.D
 Acq On : 15 Sep 2020 21:26
 Operator : JC/SP
 Sample : L3993-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAV-B38-091120-16-17

Quant Time: Sep 16 04:36:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	418194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	689802	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	688982	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	328660	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	312448	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
35) Dibromofluoromethane	5.47	113	245071	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
50) Toluene-d8	8.70	98	872551	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	11.12	95	321866	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
Target Compounds						
16) Acetone	2.42	43	27845	12.309	ug/l	99
18) Methyl Acetate	2.75	43	8851	1.621	ug/l	96
20) Methylene Chloride	2.84	84	12557	1.584	ug/l	98
43) Isopropyl Acetate	6.41	43	125396	10.506	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091520\
Data File : VX018423.D
Acq On : 15 Sep 2020 21:26
Operator : JC/SP
Sample : L3993-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B38-091120-16-17

Quant Time: Sep 16 04:36:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

