

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091520\
 Data File : VX018426.D
 Acq On : 15 Sep 2020 22:34
 Operator : JC/SP
 Sample : L4013-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAV-B40-091420-14-15

Quant Time: Sep 16 04:44:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	423503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	698783	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	709340	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	337934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	323553	55.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.06%	
35) Dibromofluoromethane	5.47	113	250999	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
50) Toluene-d8	8.70	98	889084	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	11.12	95	331178	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
Target Compounds						
16) Acetone	2.42	43	20565	8.977	ug/l	95
18) Methyl Acetate	2.75	43	7923	1.433	ug/l #	87
20) Methylene Chloride	2.84	84	18990	2.955	ug/l	92
43) Isopropyl Acetate	6.42	43	127791	10.569	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091520\
Data File : VX018426.D
Acq On : 15 Sep 2020 22:34
Operator : JC/SP
Sample : L4013-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B40-091420-14-15

Quant Time: Sep 16 04:44:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

