


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091520\
 Data File : VX018400.D
 Acq On : 15 Sep 2020 12:11
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Title : SW846 8260
 Last Update : Fri Aug 21 03:03:56 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	21624	PASS
75	95	30	60	51.6	61225	PASS
95	95	100	100	100.0	118653	PASS
96	95	5	9	6.5	7724	PASS
173	174	0.00	2	1.2	1090	PASS
174	95	50	100	79.3	94146	PASS
175	174	5	9	7.8	7348	PASS
176	174	95	101	98.3	92586	PASS
177	176	5	9	6.8	6312	PASS