

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
 Data File : VX024255.D
 Acq On : 15 Sep 2021 13:04
 Operator : JC/MD
 Sample : M3724-19 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33135

Quant Time: Sep 16 04:42:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	130816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221600	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	215904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96689	51.833	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	103.660%	
35) Dibromofluoromethane	5.397	113	74507	50.134	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.260%	
50) Toluene-d8	8.653	98	276336	51.041	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	102.080%	
62) 4-Bromofluorobenzene	11.085	95	106350	50.476	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	100.960%	
Target Compounds						Qvalue
20) Methylene Chloride	2.794	84	2068	1.354	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
Data File : VX024255.D
Acq On : 15 Sep 2021 13:04
Operator : JC/MD
Sample : M3724-19 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33135

Quant Time: Sep 16 04:42:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

