

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
 Data File : VX024257.D
 Acq On : 15 Sep 2021 13:52
 Operator : JC/MD
 Sample : M3708-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 16 04:43:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243518	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92243	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	104176	51.793	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.580%
35) Dibromofluoromethane	5.397	113	81541	49.929	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.860%
50) Toluene-d8	8.653	98	295730	49.707	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.085	95	102228	44.153	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
Data File : VX024257.D
Acq On : 15 Sep 2021 13:52
Operator : JC/MD
Sample : M3708-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 16 04:43:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

