

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
 Data File : VX024259.D
 Acq On : 15 Sep 2021 14:39
 Operator : JC/MD
 Sample : M3708-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 16 04:43:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	225633	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	192568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76740	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	99098	53.460	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.920%	
35) Dibromofluoromethane	5.397	113	76130	50.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.620%	
50) Toluene-d8	8.653	98	279126	50.635	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.280%	
62) 4-Bromofluorobenzene	11.085	95	88713	41.353	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 82.700%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
Data File : VX024259.D
Acq On : 15 Sep 2021 14:39
Operator : JC/MD
Sample : M3708-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 16 04:43:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

