

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
 Data File : VX024263.D
 Acq On : 15 Sep 2021 16:13
 Operator : JC/MD
 Sample : M3724-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33129

Quant Time: Sep 16 04:45:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105190	51.360	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.720%		
35) Dibromofluoromethane	5.397	113	82327	49.541	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.080%		
50) Toluene-d8	8.653	98	308722	50.996	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.000%		
62) 4-Bromofluorobenzene	11.085	95	118616	50.347	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.700%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	10041	11.498	ug/l	89
22) Diisopropyl ether	3.769	45	23948	5.066	ug/l #	84
31) Cyclohexane	5.476	56	12601	5.260	ug/l #	89
39) Methylcyclohexane	7.391	83	6452	2.372	ug/l #	89
68) m/p-Xylenes	10.311	106	6810	2.049	ug/l	94
69) o-Xylene	10.646	106	6214	1.873	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091521\
Data File : VX024263.D
Acq On : 15 Sep 2021 16:13
Operator : JC/MD
Sample : M3724-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33129

Quant Time: Sep 16 04:45:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

