

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031354.D
 Acq On : 15 Sep 2022 15:25
 Operator : JC/MD
 Sample : VX0915MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0915MBL01

Quant Time: Sep 16 08:02:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96073	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	161185	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95823	40.635	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.260%
35) Dibromofluoromethane	5.385	113	79948	46.143	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.280%
50) Toluene-d8	8.647	98	300871	46.293	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.580%
62) 4-Bromofluorobenzene	11.079	95	97887	41.129	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.260%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031354.D
Acq On : 15 Sep 2022 15:25
Operator : JC/MD
Sample : VX0915MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0915MBL01

Quant Time: Sep 16 08:02:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

