

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031364.D
 Acq On : 15 Sep 2022 19:19
 Operator : JC/MD
 Sample : N4492-20RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-02-133-143-090122RE

Quant Time: Sep 16 04:57:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	97711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	160645	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100199	41.778	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.560%
35) Dibromofluoromethane	5.391	113	80738	46.755	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.520%
50) Toluene-d8	8.653	98	281154	43.404	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.800%
62) 4-Bromofluorobenzene	11.079	95	89133	37.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.160%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031364.D
Acq On : 15 Sep 2022 19:19
Operator : JC/MD
Sample : N4492-20RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-02-133-143-090122RE

Quant Time: Sep 16 04:57:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

