

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031368.D
 Acq On : 15 Sep 2022 20:53
 Operator : JC/MD
 Sample : N4508-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-090222

Quant Time: Sep 16 04:58:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151847	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94014	42.060	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.120%		
35) Dibromofluoromethane	5.385	113	77767	47.644	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.280%		
50) Toluene-d8	8.653	98	297172	48.535	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.080%		
62) 4-Bromofluorobenzene	11.079	95	93595	41.744	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	83.480%		
Target Compounds						
16) Acetone	2.386	43	3105	3.254	ug/l #	82
17) Carbon Disulfide	2.514	76	799	1.138	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031368.D
Acq On : 15 Sep 2022 20:53
Operator : JC/MD
Sample : N4508-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-090222

Quant Time: Sep 16 04:58:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

