

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031369.D
 Acq On : 15 Sep 2022 21:16
 Operator : JC/MD
 Sample : N4508-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-090222

Quant Time: Sep 16 04:58:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	92926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	156386	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60757	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98121	43.018	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.040%
35) Dibromofluoromethane	5.385	113	81337	48.385	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	8.653	98	306356	48.583	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	95646	41.421	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.840%#
Target Compounds						
16) Acetone	2.386	43	6999	7.188	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031369.D
Acq On : 15 Sep 2022 21:16
Operator : JC/MD
Sample : N4508-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-090222

Quant Time: Sep 16 04:58:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

