

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031370.D
 Acq On : 15 Sep 2022 21:39
 Operator : JC/MD
 Sample : N4508-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-03-117-137-090222-H

Quant Time: Sep 16 04:58:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	152800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96827	43.101	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.200%		
35) Dibromofluoromethane	5.391	113	81101	49.377	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.760%		
50) Toluene-d8	8.653	98	307025	49.832	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.660%		
62) 4-Bromofluorobenzene	11.079	95	95965	42.534	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	85.060%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.319	96	30449	23.829	ug/l	93
16) Acetone	2.386	43	39025	40.692	ug/l	99
24) 1,1-Dichloroethane	3.617	63	6961	2.245	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	27043	14.368	ug/l	80
42) 1,2-Dichloroethane	6.092	62	4473	1.718	ug/l	96
44) Trichloroethene	7.129	130	107742	65.458	ug/l	93
49) 1,4-Dioxane	7.684	88	459	9.970	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031370.D
Acq On : 15 Sep 2022 21:39
Operator : JC/MD
Sample : N4508-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-03-117-137-090222-H

Quant Time: Sep 16 04:58:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

