

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031371.D
 Acq On : 15 Sep 2022 22:03
 Operator : JC/MD
 Sample : N4508-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-090222

Quant Time: Sep 16 04:59:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	92202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	154523	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95173	42.054	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.100%
35) Dibromofluoromethane	5.385	113	77983	46.949	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.900%
50) Toluene-d8	8.653	98	297352	47.724	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.440%
62) 4-Bromofluorobenzene	11.079	95	93544	40.999	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.000%#
Target Compounds						
16) Acetone	2.386	43	1473	1.525	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031371.D
Acq On : 15 Sep 2022 22:03
Operator : JC/MD
Sample : N4508-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-090222

Quant Time: Sep 16 04:59:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

