

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031373.D
 Acq On : 15 Sep 2022 22:50
 Operator : JC/MD
 Sample : N4508-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-090222

Quant Time: Sep 16 04:59:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	153141	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93942	41.700	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.400%
35) Dibromofluoromethane	5.391	113	78025	47.398	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.653	98	295768	47.898	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.800%
62) 4-Bromofluorobenzene	11.079	95	92421	40.872	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.740%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031373.D
Acq On : 15 Sep 2022 22:50
Operator : JC/MD
Sample : N4508-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-090222

Quant Time: Sep 16 04:59:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

