

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031379.D
 Acq On : 16 Sep 2022 02:19
 Operator : JC/MD
 Sample : VX0915WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0915WBL02

Quant Time: Sep 16 05:02:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	50732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	81216	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	86811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52962	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58529	47.002	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.000%
35) Dibromofluoromethane	5.391	113	46439	53.194	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.380%
50) Toluene-d8	8.653	98	172045	52.536	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.080%
62) 4-Bromofluorobenzene	11.079	95	53901	44.947	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031379.D
Acq On : 16 Sep 2022 02:19
Operator : JC/MD
Sample : VX0915WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0915WBL02

Quant Time: Sep 16 05:02:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

