

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031384.D
 Acq On : 16 Sep 2022 04:15
 Operator : JC/MD
 Sample : N4626-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-298-300

Quant Time: Sep 16 08:47:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	154108	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65299	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93791	41.635	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.280%
35) Dibromofluoromethane	5.391	113	76081	45.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.860%
50) Toluene-d8	8.653	98	290491	46.748	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.500%
62) 4-Bromofluorobenzene	11.079	95	91090	40.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.060%#
Target Compounds						
16) Acetone	2.392	43	3792	3.943	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031384.D
Acq On : 16 Sep 2022 04:15
Operator : JC/MD
Sample : N4626-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-298-300

Quant Time: Sep 16 08:47:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

