

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031389.D
 Acq On : 16 Sep 2022 06:12
 Operator : JC/MD
 Sample : N4626-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-408-410

Quant Time: Sep 16 08:48:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	88388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	147378	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92092	42.448	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.900%
35) Dibromofluoromethane	5.391	113	75332	47.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.653	98	281033	47.291	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.580%
62) 4-Bromofluorobenzene	11.079	95	86333	39.673	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.340%#
Target Compounds						
16) Acetone	2.386	43	2122	2.291	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031389.D
Acq On : 16 Sep 2022 06:12
Operator : JC/MD
Sample : N4626-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-408-410

Quant Time: Sep 16 08:48:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

