

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
 Data File : VX031391.D
 Acq On : 16 Sep 2022 06:59
 Operator : JC/MD
 Sample : N4626-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-EB-20220913

Quant Time: Sep 16 08:49:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	88449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	148848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88811	40.908	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	81.820%		
35) Dibromofluoromethane	5.391	113	72697	45.435	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.880%		
50) Toluene-d8	8.653	98	277523	46.240	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.480%		
62) 4-Bromofluorobenzene	11.079	95	86997	39.583	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	79.160%#		
Target Compounds					Qvalue	
30) Chloroform	5.099	83	2839	0.895	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091522\
Data File : VX031391.D
Acq On : 16 Sep 2022 06:59
Operator : JC/MD
Sample : N4626-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220913

Quant Time: Sep 16 08:49:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

