

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091522\  
 Data File : VX031397.D  
 Acq On : 16 Sep 2022 09:18  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 51 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 09/16/2022  
 Supervised By :Mahesh Dadoda 09/16/2022

Quant Time: Sep 16 18:56:18 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091422W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 15 01:50:07 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 51748      | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 84123      | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 95643      | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 74057      | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 54404      | 42.832   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 74 - 125 |      | Recovery = | 85.660%  |        |          |
| 35) Dibromofluoromethane     | 5.385          | 113  | 45231      | 50.020   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 100.040% |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 170510     | 50.268   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 86 - 113 |      | Recovery = | 100.540% |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 60973      | 49.087   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = | 98.180%  |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.167          | 85   | 25020      | 44.779   | ug/l   | 95       |
| 3) Chloromethane             | 1.288          | 50   | 38018      | 49.621   | ug/l   | 95       |
| 4) Vinyl Chloride            | 1.374          | 62   | 46480      | 51.407   | ug/l   | 96       |
| 5) Bromomethane              | 1.612          | 94   | 47307      | 69.770   | ug/l   | 98       |
| 6) Chloroethane              | 1.685          | 64   | 70727      | 98.546   | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 115382     | 69.617   | ug/l   | 98       |
| 8) Diethyl Ether             | 2.130          | 74   | 39678      | 60.073   | ug/l   | 87       |
| 9) 1,1,2-Trichlorotrifluo... | 2.331          | 101  | 38685      | 49.359   | ug/l   | 94       |
| 10) Methyl Iodide            | 2.453          | 142  | 31668      | 38.090   | ug/l   | 99       |
| 11) Tert butyl alcohol       | 3.008          | 59   | 63859      | 239.204  | ug/l # | 94       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 37342      | 51.685   | ug/l   | 94       |
| 13) Acrolein                 | 2.240          | 56   | 26201      | 178.754  | ug/l   | 99       |
| 14) Allyl chloride           | 2.666          | 41   | 59518      | 37.290   | ug/l   | 98       |
| 15) Acrylonitrile            | 3.069          | 53   | 127518     | 217.669  | ug/l   | 100      |
| 16) Acetone                  | 2.392          | 43   | 112064     | 206.668  | ug/l   | 98       |
| 17) Carbon Disulfide         | 2.514          | 76   | 82823      | 47.943   | ug/l   | 98       |
| 18) Methyl Acetate           | 2.709          | 43   | 66813      | 41.552   | ug/l   | 98       |
| 19) Methyl tert-butyl Ether  | 3.117          | 73   | 148172     | 44.287   | ug/l   | 100      |
| 20) Methylene Chloride       | 2.788          | 84   | 44556      | 52.516   | ug/l   | 94       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 42421      | 52.313   | ug/l   | 94       |
| 22) Diisopropyl ether        | 3.764          | 45   | 149173     | 44.664   | ug/l   | 97       |
| 23) Vinyl Acetate            | 3.727          | 43   | 641198     | 231.005  | ug/l   | 98       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 79765      | 45.507   | ug/l   | 99       |
| 25) 2-Butanone               | 4.568          | 43   | 185313     | 207.191  | ug/l   | 97       |
| 26) 2,2-Dichloropropane      | 4.477          | 77   | 42258      | 25.411   | ug/l   | 93       |
| 27) cis-1,2-Dichloroethene   | 4.489          | 96   | 52161      | 49.014   | ug/l   | 87       |
| 28) Bromochloromethane       | 4.898          | 49   | 34095      | 41.732   | ug/l   | 94       |
| 29) Tetrahydrofuran          | 5.013          | 42   | 119496     | 215.692  | ug/l   | 97       |
| 30) Chloroform               | 5.099          | 83   | 89689      | 48.321   | ug/l   | 100      |
| 31) Cyclohexane              | 5.477          | 56   | 65060      | 47.677   | ug/l   | 91       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 78947      | 47.468   | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 60797      | 49.121   | ug/l   | 99       |
| 37) Ethyl Acetate            | 4.715          | 43   | 71879      | 44.193   | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 67562      | 50.218   | ug/l   | 98       |
| 39) Methylcyclohexane        | 7.385          | 83   | 66961      | 49.126   | ug/l   | 93       |
| 40) Benzene                  | 6.044          | 78   | 184165     | 51.817   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091522\  
 Data File : VX031397.D  
 Acq On : 16 Sep 2022 09:18  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 51 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 09/16/2022  
 Supervised By :Mahesh Dadoda 09/16/2022

Quant Time: Sep 16 18:56:18 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091422W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 15 01:50:07 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.928  | 41   | 37224    | 42.246  | ug/l   | 94       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 71234    | 49.688  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 123645   | 44.536  | ug/l   | 99       |
| 44) Trichloroethene           | 7.129  | 130  | 50594    | 55.832  | ug/l   | 90       |
| 45) 1,2-Dichloropropane       | 7.434  | 63   | 47473    | 48.790  | ug/l   | 98       |
| 46) Dibromomethane            | 7.580  | 93   | 35863    | 51.973  | ug/l   | 94       |
| 47) Bromodichloromethane      | 7.824  | 83   | 70923    | 49.416  | ug/l   | 95       |
| 48) Methyl methacrylate       | 7.696  | 41   | 53635    | 42.438  | ug/l   | 92       |
| 49) 1,4-Dioxane               | 7.702  | 88   | 25346    | 999.997 | ug/l   | 95       |
| 51) 4-Methyl-2-Pentanone      | 8.574  | 43   | 403103   | 238.868 | ug/l   | 100      |
| 52) Toluene                   | 8.720  | 92   | 124118   | 55.206  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 71606    | 44.335  | ug/l   | 100      |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 73512    | 44.290  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 52774    | 53.174  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 80253    | 49.951  | ug/l   | 95       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 85781    | 50.300  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 8.245  | 63   | 169731   | 226.738 | ug/l   | 96       |
| 59) 2-Hexanone                | 9.433  | 43   | 309058   | 239.953 | ug/l   | 100      |
| 60) Dibromochloromethane      | 9.525  | 129  | 57931    | 54.142  | ug/l   | 98       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 56159    | 55.146  | ug/l   | 98       |
| 64) Tetrachloroethene         | 9.275  | 164  | 48735    | 55.731  | ug/l   | 97       |
| 65) Chlorobenzene             | 10.080 | 112  | 136107   | 53.731  | ug/l   | 100      |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 51803    | 51.905  | ug/l   | 98       |
| 67) Ethyl Benzene             | 10.195 | 91   | 240039   | 52.318  | ug/l   | 100      |
| 68) m/p-Xylenes               | 10.305 | 106  | 193494   | 111.066 | ug/l   | 99       |
| 69) o-Xylene                  | 10.640 | 106  | 95906    | 54.200  | ug/l   | 97       |
| 70) Styrene                   | 10.659 | 104  | 160534   | 55.791  | ug/l   | 99       |
| 71) Bromoform                 | 10.799 | 173  | 42261    | 53.190  | ug/l # | 100      |
| 73) Isopropylbenzene          | 10.964 | 105  | 253196   | 46.289  | ug/l   | 98       |
| 74) N-amyl acetate            | 10.842 | 43   | 111381   | 40.959  | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 11.214 | 83   | 85672    | 43.582  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 11.238 | 75   | 77114m   | 41.320  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 60217    | 48.281  | ug/l   | 97       |
| 78) n-propylbenzene           | 11.305 | 91   | 291274   | 45.374  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 170828   | 43.491  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 218034   | 47.577  | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 21391    | 31.716  | ug/l   | 96       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 212724   | 47.249  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 11.713 | 119  | 214792   | 46.846  | ug/l   | 100      |
| 84) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 214858   | 47.171  | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.890 | 105  | 270022   | 47.031  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 231028   | 49.381  | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 120030   | 50.479  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.043 | 146  | 127192   | 52.045  | ug/l   | 98       |
| 89) n-Butylbenzene            | 12.335 | 91   | 195756   | 47.185  | ug/l   | 98       |
| 90) Hexachloroethane          | 12.543 | 117  | 40331    | 43.167  | ug/l   | 97       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 122605   | 50.644  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 18347    | 40.327  | ug/l   | 90       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 68622    | 52.347  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 23116    | 45.174  | ug/l   | 96       |
| 95) Naphthalene               | 13.774 | 128  | 256031   | 48.072  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 69169    | 49.272  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091522\  
 Data File : VX031397.D  
 Acq On : 16 Sep 2022 09:18  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 51 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 09/16/2022  
 Supervised By :Mahesh Dadoda 09/16/2022

Quant Time: Sep 16 18:56:18 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091422W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 15 01:50:07 2022  
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091522\  
 Data File : VX031397.D  
 Acq On : 16 Sep 2022 09:18  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 51 Sample Multiplier: 1


Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Quant Time: Sep 16 18:56:18 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091422W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 15 01:50:07 2022  
 Response via : Initial Calibration

### Manual Integrations APPROVED


Reviewed By :Krupa Patel 09/16/2022  
 Supervised By :Mahesh Dadoda 09/16/2022





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:168 Resp: 51748  
Ion Ratio Lower Upper  
168 100  
99 53.1 48.8 73.2




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#2  
Dichlorodifluoromethane  
Concen: 44.779 ug/l  
RT: 1.167 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion: 85 Resp: 25020  
Ion Ratio Lower Upper  
85 100  
87 34.2 15.8 47.4





#3  
Chloromethane  
Concen: 49.621 ug/l  
RT: 1.288 min Scan# 33  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 50 Resp: 38018  
Ion Ratio Lower Upper  
50 100  
52 30.7 26.9 40.3

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#4  
Vinyl Chloride  
Concen: 51.407 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 62 Resp: 46480  
Ion Ratio Lower Upper  
62 100  
64 33.8 25.3 37.9





#5  
Bromomethane  
Concen: 69.770 ug/l  
RT: 1.612 min Scan# 80  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 94 Resp: 4730  
Ion Ratio Lower Upper  
94 100  
96 95.2 74.4 111.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#6  
Chloroethane  
Concen: 98.546 ug/l  
RT: 1.685 min Scan# 98  
Delta R.T. 0.007 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18



Tgt Ion: 64 Resp: 70727  
Ion Ratio Lower Upper  
64 100  
66 32.8 25.0 37.4









#7  
Trichlorofluoromethane  
Concen: 69.617 ug/l  
RT: 1.886 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion:101 Resp: 11538  
Ion Ratio Lower Upper  
101 100  
103 66.4 51.8 77.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#8  
Diethyl Ether  
Concen: 60.073 ug/l  
RT: 2.130 min Scan# 171  
Delta R.T. -0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Tgt Ion: 74 Resp: 39678  
Ion Ratio Lower Upper  
74 100  
45 84.4 48.5 145.6









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 49.359 ug/l  
RT: 2.331 min Scan# 203  
Delta R.T. 0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion:101 Resp: 3868  
Ion Ratio Lower Upper  
101 100  
85 47.4 38.6 57.8  
151 75.7 54.0 81.0




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#10  
Methyl Iodide  
Concen: 38.090 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion:142 Resp: 31668  
Ion Ratio Lower Upper  
142 100  
127 47.6 37.4 56.0  
141 14.5 11.6 17.4





#11  
Tert butyl alcohol  
Concen: 239.204 ug/l  
RT: 3.008 min Scan# 312  
Delta R.T. 0.019 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 59 Resp: 6385  
Ion Ratio Lower Upper  
59 100  
57 7.9 8.2 12.2




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#12  
1,1-Dichloroethene  
Concen: 51.685 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion: 96 Resp: 37342  
Ion Ratio Lower Upper  
96 100  
61 159.0 135.3 202.9  
98 60.3 49.4 74.2





#13  
Acrolein  
Concen: 178.754 ug/l  
RT: 2.240 min Scan# 189  
Delta R.T. 0.001 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 56 Resp: 2620  
Ion Ratio Lower Upper  
56 100  
55 69.0 56.0 84.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#14  
Allyl chloride  
Concen: 37.290 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 41 Resp: 59518  
Ion Ratio Lower Upper  
41 100  
39 73.3 58.2 87.4  
76 36.5 26.9 40.3





#15  
Acrylonitrile  
Concen: 217.669 ug/l  
RT: 3.069 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 53 Resp: 127518  
Ion Ratio Lower Upper  
53 100  
52 83.2 66.9 100.3  
51 37.2 30.0 45.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#16  
Acetone  
Concen: 206.668 ug/l  
RT: 2.392 min Scan# 214  
Delta R.T. 0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 43 Resp: 112064  
Ion Ratio Lower Upper  
43 100  
58 31.1 23.8 35.8





#17  
Carbon Disulfide  
Concen: 47.943 ug/l  
RT: 2.514 min Scan# 21  
Delta R.T. 0.007 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 76 Resp: 8282  
Ion Ratio Lower Upper  
76 100  
78 9.6 7.2 10.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#18  
Methyl Acetate  
Concen: 41.552 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. 0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 43 Resp: 66813  
Ion Ratio Lower Upper  
43 100  
74 24.0 18.6 27.8





#19  
Methyl tert-butyl Ether  
Concen: 44.287 ug/l  
RT: 3.117 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 73 Resp: 148172  
Ion Ratio Lower Upper  
73 100  
57 21.6 17.2 25.8




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#20  
Methylene Chloride  
Concen: 52.516 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion: 84 Resp: 44556  
Ion Ratio Lower Upper  
84 100  
49 122.7 105.9 158.9  
51 36.9 32.2 48.2  
86 66.0 50.8 76.2





#21  
trans-1,2-Dichloroethene  
Concen: 52.313 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 96 Resp: 4242  
Ion Ratio Lower Upper  
96 100  
61 124.4 106.7 160.1  
98 62.4 48.7 73.1




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#22  
Diisopropyl ether  
Concen: 44.664 ug/l  
RT: 3.764 min Scan# 439  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion: 45 Resp: 149173  
Ion Ratio Lower Upper  
45 100  
43 77.5 59.0 88.6  
87 29.1 22.6 33.8  
59 11.2 8.8 13.2





#23  
 Vinyl Acetate  
 Concen: 231.005 ug/l  
 RT: 3.727 min Scan# 411  
 Delta R.T. 0.006 min  
 Lab File: VX031397.D  
 Acq: 16 Sep 2022 09:18




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC




Tgt Ion: 43 Resp: 641198  
 Ion Ratio Lower Upper  
 43 100  
 86 11.7 8.7 13.1

Manual Integrations  
 APPROVED



Reviewed By :Krupa Patel 09/16/2022  
 Supervised By :Mahesh Dadoda 09/16/2022




#24  
 1,1-Dichloroethane  
 Concen: 45.507 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. 0.000 min  
 Lab File: VX031397.D  
 Acq: 16 Sep 2022 09:18




Tgt Ion: 63 Resp: 79765  
 Ion Ratio Lower Upper  
 63 100  
 98 7.3 3.5 10.4  
 100 5.3 2.4 7.2





#25  
2-Butanone  
Concen: 207.191 ug/l  
RT: 4.568 min Scan# 570  
Delta R.T. 0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 18531  
Ion Ratio Lower Upper  
43 100  
72 27.7 21.1 31.7




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#26  
2,2-Dichloropropane  
Concen: 25.411 ug/l  
RT: 4.477 min Scan# 556  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion: 77 Resp: 42258  
Ion Ratio Lower Upper  
77 100  
97 24.9 10.9 32.6





#27  
 cis-1,2-Dichloroethene  
 Concen: 49.014 ug/l  
 RT: 4.489 min Scan# 516  
 Delta R.T. 0.000 min  
 Lab File: VX031397.D  
 Acq: 16 Sep 2022 09:18




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC




Tgt Ion: 96 Resp: 5216  
 Ion Ratio Lower Upper  
 96 100  
 61 129.6 0.0 305.2  
 98 63.7 0.0 128.4

Manual Integrations  
 APPROVED



Reviewed By :Krupa Patel 09/16/2022  
 Supervised By :Mahesh Dadoda 09/16/2022




#28  
 Bromochloromethane  
 Concen: 41.732 ug/l  
 RT: 4.898 min Scan# 625  
 Delta R.T. 0.000 min  
 Lab File: VX031397.D  
 Acq: 16 Sep 2022 09:18



Tgt Ion: 49 Resp: 34095  
 Ion Ratio Lower Upper  
 49 100  
 129 2.1 0.0 3.6  
 130 78.9 59.0 88.4





#29

Tetrahydrofuran

Concen: 215.692 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX031397.D


Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 42    | Resp: | 119490 |
| Ion      | Ratio | Lower | Upper  |
| 42       | 100   |       |        |
| 72       | 47.2  | 36.4  | 54.6   |
| 71       | 44.1  | 33.5  | 50.3   |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#30

Chloroform

Concen: 48.321 ug/l




RT: 5.099 min Scan# 658


Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 89689 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 85       | 66.3  | 52.8  | 79.2  |





#31  
Cyclohexane  
Concen: 47.677 ug/l  
RT: 5.477 min Scan# 719  
Delta R.T. 0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 56 Resp: 65060  
Ion Ratio Lower Upper  
56 100  
69 31.4 25.1 37.7  
84 98.5 70.2 105.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#32  
1,1,1-Trichloroethane  
Concen: 47.468 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18



Tgt Ion: 97 Resp: 78947  
Ion Ratio Lower Upper  
97 100  
99 64.4 51.3 76.9  
61 45.7 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 42.832 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX031397.D


Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 65 Resp: 5440

Ion Ratio Lower Upper

65 100




67 52.3 0.0 102.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18




Tgt Ion:114 Resp: 84123


Ion Ratio Lower Upper

114 100

63 20.6 0.0 42.6

88 14.8 0.0 33.0





#35

Dibromofluoromethane

Concen: 50.020 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC





|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 113   | Resp: | 4523  |
| Ion Ratio | Lower | Upper |       |
| 113       | 100   |       |       |
| 111       | 101.0 | 82.4  | 123.6 |
| 192       | 18.4  | 14.1  | 21.1  |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#36

1,1-Dichloropropene

Concen: 49.121 ug/l




RT: 5.696 min Scan# 756


Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 75    | Resp: | 60797 |
| Ion Ratio | Lower | Upper |       |
| 75        | 100   |       |       |
| 110       | 36.1  | 17.5  | 52.5  |
| 77        | 31.0  | 24.9  | 37.3  |






#37  
Ethyl Acetate  
Concen: 44.193 ug/l  
RT: 4.715 min Scan# 596  
Delta R.T. -0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 43 Resp: 71879  
Ion Ratio Lower Upper  
43 100  
61 14.2 11.1 16.7  
70 10.1 8.6 13.0

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 09/16/2022


Supervised By :Mahesh Dadoda 09/16/2022



#38  
Carbon Tetrachloride  
Concen: 50.218 ug/l  
RT: 5.678 min Scan# 753  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Tgt Ion:117 Resp: 67562  
Ion Ratio Lower Upper  
117 100  
119 94.4 74.6 112.0  
121 29.4 25.3 37.9





#39

Methylcyclohexane

Concen: 49.126 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.006 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC





|             |             |
|-------------|-------------|
| Tgt Ion: 83 | Resp: 6696  |
| Ion Ratio   | Lower Upper |
| 83 100      |             |
| 55 68.9     | 61.0 91.4   |
| 98 49.6     | 37.7 56.5   |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#40

Benzene

Concen: 51.817 ug/l




RT: 6.044 min Scan# 813


Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18


|             |              |
|-------------|--------------|
| Tgt Ion: 78 | Resp: 184165 |
| Ion Ratio   | Lower Upper  |
| 78 100      |              |
| 77 23.9     | 18.7 28.1    |





#41  
Methacrylonitrile  
Concen: 42.246 ug/l  
RT: 4.928 min Scan# 611  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




| Tgt Ion: | 41    | Resp: | 37224 |
|----------|-------|-------|-------|
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 59.9  | 45.0  | 67.6  |
| 67       | 78.6  | 58.2  | 87.4  |
| 52       | 34.2  | 25.0  | 37.6  |

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#42  
1,2-Dichloroethane  
Concen: 49.688 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




| Tgt Ion: | 62    | Resp: | 71234 |
|----------|-------|-------|-------|
| Ion      | Ratio | Lower | Upper |
| 62       | 100   |       |       |
| 98       | 9.4   | 0.0   | 17.8  |





#43  
Isopropyl Acetate  
Concen: 44.536 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 123649  
Ion Ratio Lower Upper  
43 100  
61 18.6 15.3 22.9  
87 13.1 10.6 15.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#44  
Trichloroethene  
Concen: 55.832 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Tgt Ion:130 Resp: 50594  
Ion Ratio Lower Upper  
130 100  
95 93.3 0.0 206.8





#45

1,2-Dichloropropane

Concen: 48.790 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 4747

Ion Ratio Lower Upper

63 100





65 32.0 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#46

Dibromomethane

Concen: 51.973 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18




Tgt Ion: 93 Resp: 35863

Ion Ratio Lower Upper

93 100

95 81.3 65.5 98.3


174 94.1 67.6 101.4





#47  
Bromodichloromethane  
Concen: 49.416 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 70923  
Ion Ratio Lower Upper  
83 100  
85 68.3 51.0 76.6  
127 9.7 7.0 10.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#48  
Methyl methacrylate  
Concen: 42.438 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18



Tgt Ion: 41 Resp: 53635  
Ion Ratio Lower Upper  
41 100  
69 86.0 64.6 96.8  
39 72.9 51.4 77.0





#49

1,4-Dioxane

Concen: 999.997 ug/l

RT: 7.702 min Scan# 1082

Delta R.T. 0.019 min

Lab File: VX031397.D


Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 88    | Resp: | 25340 |
| Ion      | Ratio | Lower | Upper |
| 88       | 100   |       |       |
| 43       | 33.3  | 27.1  | 40.7  |
| 58       | 61.5  | 53.9  | 80.9  |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#50

Toluene-d8

Concen: 50.268 ug/l




RT: 8.653 min Scan# 1241

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 98    | Resp: | 170510 |
| Ion      | Ratio | Lower | Upper  |
| 98       | 100   |       |        |
| 100      | 65.0  | 52.5  | 78.7   |





#51  
4-Methyl-2-Pentanone  
Concen: 238.868 ug/l  
RT: 8.574 min Scan# 1228  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 40310  
Ion Ratio Lower Upper  
43 100  
58 38.1 30.6 46.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#52  
Toluene  
Concen: 55.206 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 92 Resp: 124118  
Ion Ratio Lower Upper  
92 100  
91 171.5 137.1 205.7





#53  
t-1,3-Dichloropropene  
Concen: 44.335 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 75 Resp: 71600  
Ion Ratio Lower Upper  
75 100  
77 31.1 24.8 37.2




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#54  
cis-1,3-Dichloropropene  
Concen: 44.290 ug/l  
RT: 8.366 min Scan# 1194  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Tgt Ion: 75 Resp: 73512  
Ion Ratio Lower Upper  
75 100  
77 32.6 24.4 36.6  
39 55.8 43.7 65.5





#55

1,1,2-Trichloroethane

Concen: 53.174 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC





|             |             |
|-------------|-------------|
| Tgt Ion: 97 | Resp: 52774 |
| Ion Ratio   | Lower Upper |
| 97 100      |             |
| 83 87.6     | 69.0 103.6  |
| 85 56.5     | 44.1 66.1   |
| 99 61.4     | 51.3 76.9   |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#56

Ethyl methacrylate

Concen: 49.951 ug/l




RT: 9.116 min Scan# 1317


Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

|             |             |
|-------------|-------------|
| Tgt Ion: 69 | Resp: 80253 |
| Ion Ratio   | Lower Upper |
| 69 100      |             |
| 41 65.4     | 54.2 81.4   |
| 39 48.2     | 34.2 51.2   |





#57

1,3-Dichloropropane

Concen: 50.300 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.000 min

Lab File: VX031397.D


Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 76 Resp: 8578

Ion Ratio Lower Upper

76 100



78 32.8 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#58

2-Chloroethyl Vinyl ether

Concen: 226.738 ug/l

RT: 8.245 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX031397.D




Acq: 16 Sep 2022 09:18


Tgt Ion: 63 Resp: 169731

Ion Ratio Lower Upper

63 100

106 30.2 22.6 34.0









#59  
2-Hexanone  
Concen: 239.953 ug/l  
RT: 9.433 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 43 Resp: 309058  
Ion Ratio Lower Upper  
43 100  
58 53.2 26.7 80.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#60  
Dibromochloromethane  
Concen: 54.142 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion:129 Resp: 57931  
Ion Ratio Lower Upper  
129 100  
127 75.3 38.7 116.1





#61  
1,2-Dibromoethane  
Concen: 55.146 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:107 Resp: 56159  
Ion Ratio Lower Upper  
107 100  
109 94.5 74.2 111.4




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#62  
4-Bromofluorobenzene  
Concen: 49.087 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion: 95 Resp: 60973  
Ion Ratio Lower Upper  
95 100  
174 84.2 0.0 153.2  
176 81.8 0.0 144.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:117 Resp: 9564  
Ion Ratio Lower Upper  
117 100  
82 50.3 46.8 70.2  
119 31.4 26.4 39.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#64  
Tetrachloroethene  
Concen: 55.731 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion:164 Resp: 48735  
Ion Ratio Lower Upper  
164 100  
166 130.8 99.0 148.4  
129 96.4 78.0 117.0  
131 92.3 74.2 111.2





#65  
Chlorobenzene  
Concen: 53.731 ug/l  
RT: 10.080 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:112 Resp: 13610  
Ion Ratio Lower Upper  
112 100  
114 31.8 25.7 38.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 51.905 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18



Tgt Ion:131 Resp: 51803  
Ion Ratio Lower Upper  
131 100  
133 95.2 46.6 139.7  
119 65.3 31.9 95.9





#67

Ethyl Benzene

Concen: 52.318 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 240039

Ion Ratio Lower Upper

91 100





106 31.3 25.2 37.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#68

m/p-Xylenes

Concen: 111.066 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX031397.D




Acq: 16 Sep 2022 09:18


Tgt Ion:106 Resp: 193494

Ion Ratio Lower Upper

106 100


91 199.1 161.0 241.4





#69  
o-Xylene  
Concen: 54.200 ug/l  
RT: 10.640 min Scan# 1568  
Delta R.T. -0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:106 Resp: 95900  
Ion Ratio Lower Upper  
106 100  
91 210.4 107.2 321.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#70  
Styrene  
Concen: 55.791 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18


Tgt Ion:104 Resp: 160534  
Ion Ratio Lower Upper  
104 100  
78 50.1 41.4 62.2  
103 54.7 43.8 65.8





#71  
Bromoform  
Concen: 53.190 ug/l  
RT: 10.799 min Scan# 1593  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:173 Resp: 4226  
Ion Ratio Lower Upper  
173 100  
175 47.9 24.0 72.0  
254 0.0 0.0 0.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion:152 Resp: 74057  
Ion Ratio Lower Upper  
152 100  
115 84.1 43.4 130.1  
150 173.6 0.0 346.8





#73  
Isopropylbenzene  
Concen: 46.289 ug/l  
RT: 10.964 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 105 Resp: 253190  
Ion Ratio Lower Upper  
105 100  
120 26.4 13.7 41.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#74  
N-ethyl acetate  
Concen: 40.959 ug/l  
RT: 10.842 min Scan# 1600  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 43 Resp: 111381  
Ion Ratio Lower Upper  
43 100  
70 41.4 33.4 50.2  
55 22.3 18.2 27.2  
61 23.0 19.3 28.9





#75  
1,1,2,2-Tetrachloroethane  
Concen: 43.582 ug/l  
RT: 11.214 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 85672  
Ion Ratio Lower Upper  
83 100  
131 9.7 5.1 15.4  
85 63.3 32.6 97.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#76  
1,2,3-Trichloropropane  
Concen: 41.320 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18



Tgt Ion: 75 Resp: 77114  
Ion Ratio Lower Upper  
75 100  
77 37.3 20.3 60.9





#77

Bromobenzene

Concen: 48.281 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC





|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 156   | Resp: | 6021  |
| Ion Ratio | Lower | Upper |       |
| 156       | 100   |       |       |
| 77        | 152.8 | 79.5  | 238.6 |
| 158       | 97.5  | 48.4  | 145.3 |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#78

n-propylbenzene

Concen: 45.374 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 91    | Resp: | 291274 |
| Ion Ratio | Lower | Upper |        |
| 91        | 100   |       |        |
| 120       | 23.5  | 11.8  | 35.4   |





#79  
2-Chlorotoluene  
Concen: 43.491 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 91 Resp: 170828  
Ion Ratio Lower Upper  
91 100  
126 34.5 17.4 52.3

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#80  
1,3,5-Trimethylbenzene  
Concen: 47.577 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion:105 Resp: 218034  
Ion Ratio Lower Upper  
105 100  
120 50.2 24.5 73.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 31.716 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 75 Resp: 2139  
Ion Ratio Lower Upper  
75 100  
53 103.0 77.6 116.4  
89 47.3 37.4 56.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#82  
4-Chlorotoluene  
Concen: 47.249 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 91 Resp: 212724  
Ion Ratio Lower Upper  
91 100  
126 30.0 15.1 45.3





#83  
tert-Butylbenzene  
Concen: 46.846 ug/l  
RT: 11.713 min Scan# 11  
Delta R.T. -0.006 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:119 Resp: 214792  
Ion Ratio Lower Upper  
119 100  
91 58.4 29.2 87.6  
134 23.4 11.6 34.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#84  
1,2,4-Trimethylbenzene  
Concen: 47.171 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion:105 Resp: 214858  
Ion Ratio Lower Upper  
105 100  
120 45.5 22.6 67.8





#85  
 sec-Butylbenzene  
 Concen: 47.031 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. 0.000 min  
 Lab File: VX031397.D  
 Acq: 16 Sep 2022 09:18




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC




Tgt Ion:105 Resp: 270022  
 Ion Ratio Lower Upper  
 105 100  
 134 20.2 10.4 31.1

Manual Integrations  
 APPROVED



Reviewed By :Krupa Patel 09/16/2022  
 Supervised By :Mahesh Dadoda 09/16/2022




#86  
 p-Isopropyltoluene  
 Concen: 49.381 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. 0.000 min  
 Lab File: VX031397.D  
 Acq: 16 Sep 2022 09:18



Tgt Ion:119 Resp: 231028  
 Ion Ratio Lower Upper  
 119 100  
 134 26.8 13.4 40.2  
 91 25.0 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 50.479 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 120030

Ion Ratio Lower Upper

146 100

111 39.4 19.8 59.4





148 63.9 31.5 94.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/16/2022

Supervised By :Mahesh Dadoda 09/16/2022



#88

1,4-Dichlorobenzene

Concen: 52.045 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX031397.D

Acq: 16 Sep 2022 09:18




Tgt Ion:146 Resp: 127192

Ion Ratio Lower Upper

146 100

111 38.8 20.5 61.5


148 64.3 32.0 96.2





#89  
n-Butylbenzene  
Concen: 47.185 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 91 Resp: 195750  
Ion Ratio Lower Upper  
91 100  
92 52.1 27.0 81.0  
134 25.6 13.3 39.9

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#90  
Hexachloroethane  
Concen: 43.167 ug/l  
RT: 12.543 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion:117 Resp: 40331  
Ion Ratio Lower Upper  
117 100  
201 67.1 34.8 104.4





#91  
1,2-Dichlorobenzene  
Concen: 50.644 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:146 Resp: 122609  
Ion Ratio Lower Upper  
146 100  
111 40.0 20.6 61.8  
148 63.9 31.6 95.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 40.327 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion: 75 Resp: 18347  
Ion Ratio Lower Upper  
75 100  
155 89.0 40.3 120.8  
157 113.9 51.5 154.5





#93  
1,2,4-Trichlorobenzene  
Concen: 52.347 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:180 Resp: 6862  
Ion Ratio Lower Upper  
180 100  
182 93.6 46.9 140.6  
145 29.9 16.7 50.1

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022




#94  
Hexachlorobutadiene  
Concen: 45.174 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Tgt Ion:225 Resp: 23116  
Ion Ratio Lower Upper  
225 100  
223 64.2 30.8 92.4  
227 60.7 32.3 96.9





#95  
Naphthalene  
Concen: 48.072 ug/l  
RT: 13.774 min Scan# 2081  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:128 Resp: 25603  
Ion Ratio Lower Upper  
128 100  
127 13.0 10.4 15.6  
129 11.1 8.8 13.2

Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 09/16/2022  
Supervised By :Mahesh Dadoda 09/16/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 49.272 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. 0.000 min  
Lab File: VX031397.D  
Acq: 16 Sep 2022 09:18

Tgt Ion:180 Resp: 69169  
Ion Ratio Lower Upper  
180 100  
182 94.8 47.9 143.8  
145 32.9 17.6 52.8

