

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
 Data File : VX037746.D
 Acq On : 15 Sep 2023 10:13
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Sep 16 04:29:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:24:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	259344	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	430033	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	410931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	205192	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	170017	49.897	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.800%
35) Dibromofluoromethane	5.385	113	145313	50.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.000%
50) Toluene-d8	8.647	98	546543	51.121	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.240%
62) 4-Bromofluorobenzene	11.079	95	196656	50.619	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.240%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.166	85	134265	49.117	ug/l	99
3) Chloromethane	1.300	50	137461	46.217	ug/l	97
4) Vinyl Chloride	1.374	62	157573	49.108	ug/l	98
5) Bromomethane	1.599	94	115789	50.406	ug/l	97
6) Chloroethane	1.678	64	106087	47.992	ug/l	99
7) Trichlorofluoromethane	1.886	101	259698	50.343	ug/l	99
8) Diethyl Ether	2.136	74	96376	48.223	ug/l	82
9) 1,1,2-Trichlorotrifluo...	2.331	101	130109	49.699	ug/l	97
10) Methyl Iodide	2.453	142	179570	51.831	ug/l	93
11) Tert butyl alcohol	2.965	59	176968	241.962	ug/l	# 92
12) 1,1-Dichloroethene	2.319	96	126487	49.992	ug/l	86
13) Acrolein	2.239	56	175380	242.726	ug/l	99
14) Allyl chloride	2.666	41	207806	49.314	ug/l	# 95
15) Acrylonitrile	3.062	53	404033	248.462	ug/l	99
16) Acetone	2.380	43	375160	228.873	ug/l	# 89
17) Carbon Disulfide	2.514	76	316209	49.847	ug/l	99
18) Methyl Acetate	2.703	43	289189	48.906	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	455989	48.992	ug/l	100
20) Methylene Chloride	2.788	84	150172	49.554	ug/l	90
21) trans-1,2-Dichloroethene	3.093	96	140549	50.026	ug/l	90
22) Diisopropyl ether	3.763	45	437172	49.805	ug/l	# 85
23) Vinyl Acetate	3.721	43	1646586	255.031	ug/l	95
24) 1,1-Dichloroethane	3.611	63	257515	48.958	ug/l	99
25) 2-Butanone	4.556	43	567078	245.222	ug/l	# 90
26) 2,2-Dichloropropane	4.477	77	209137	49.207	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	171445	49.776	ug/l	89
28) Bromochloromethane	4.897	49	117007	47.683	ug/l	# 83
29) Tetrahydrofuran	5.007	42	358180	248.957	ug/l	89
30) Chloroform	5.092	83	280064	49.427	ug/l	97
31) Cyclohexane	5.470	56	222993	50.989	ug/l	95
32) 1,1,1-Trichloroethane	5.385	97	245770	51.226	ug/l	97
36) 1,1-Dichloropropene	5.696	75	197883	49.923	ug/l	96
37) Ethyl Acetate	4.714	43	217161	51.031	ug/l	97
38) Carbon Tetrachloride	5.678	117	212170	51.159	ug/l	97
39) Methylcyclohexane	7.379	83	253272	50.346	ug/l	94
40) Benzene	6.037	78	605328	49.662	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
 Data File : VX037746.D
 Acq On : 15 Sep 2023 10:13
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Sep 16 04:29:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:24:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	110678	49.813	ug/l #	88
42) 1,2-Dichloroethane	6.086	62	215577	48.785	ug/l	96
43) Isopropyl Acetate	6.342	43	353314	50.707	ug/l	96
44) Trichloroethene	7.129	130	161598	47.606	ug/l	97
45) 1,2-Dichloropropane	7.433	63	155431	49.103	ug/l	99
46) Dibromomethane	7.580	93	114355	50.222	ug/l	93
47) Bromodichloromethane	7.824	83	216033	51.515	ug/l	97
48) Methyl methacrylate	7.696	41	169547	50.902	ug/l #	87
49) 1,4-Dioxane	7.659	88	84402	966.542	ug/l #	89
51) 4-Methyl-2-Pentanone	8.574	43	1114690	252.149	ug/l	94
52) Toluene	8.720	92	398453	50.723	ug/l	96
53) t-1,3-Dichloropropene	8.976	75	232974	53.372	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	251705	51.867	ug/l #	89
55) 1,1,2-Trichloroethane	9.153	97	167342	50.526	ug/l	95
56) Ethyl methacrylate	9.116	69	255142	51.602	ug/l	91
57) 1,3-Dichloropropane	9.311	76	266525	49.730	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	568955	246.430	ug/l	99
59) 2-Hexanone	9.433	43	864988	251.303	ug/l	94
60) Dibromochloromethane	9.525	129	174770	52.971	ug/l	100
61) 1,2-Dibromoethane	9.610	107	179632	51.226	ug/l	99
64) Tetrachloroethene	9.275	164	135780	48.450	ug/l	96
65) Chlorobenzene	10.079	112	429927	48.504	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	162476	51.725	ug/l	99
67) Ethyl Benzene	10.195	91	769315	50.061	ug/l	99
68) m/p-Xylenes	10.299	106	600485	101.121	ug/l	97
69) o-Xylene	10.640	106	299129	50.233	ug/l	97
70) Styrene	10.652	104	494103	51.686	ug/l	96
71) Bromoform	10.799	173	130939	52.834	ug/l #	100
73) Isopropylbenzene	10.963	105	776441	51.577	ug/l	98
74) N-amyl acetate	10.841	43	319637	52.428	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	262045	49.495	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	287462	57.246	ug/l	90
77) Bromobenzene	11.195	156	187998	50.455	ug/l	89
78) n-propylbenzene	11.305	91	898303	52.838	ug/l	99
79) 2-Chlorotoluene	11.366	91	531369	50.690	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	664087	52.237	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	79232	53.480	ug/l	89
82) 4-Chlorotoluene	11.457	91	608183	51.534	ug/l	96
83) tert-Butylbenzene	11.713	119	671758	52.282	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	658015	52.022	ug/l	98
85) sec-Butylbenzene	11.890	105	822306	52.752	ug/l	99
86) p-Isopropyltoluene	12.012	119	681771	52.579	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	345898	49.801	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	344312	48.586	ug/l	97
89) n-Butylbenzene	12.335	91	604859	53.164	ug/l	97
90) Hexachloroethane	12.536	117	117268	53.292	ug/l	88
91) 1,2-Dichlorobenzene	12.335	146	346286	48.832	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	55489	52.579	ug/l	77
93) 1,2,4-Trichlorobenzene	13.585	180	213559	51.125	ug/l	98
94) Hexachlorobutadiene	13.725	225	99452	51.017	ug/l	98
95) Naphthalene	13.774	128	738377	52.266	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	217481	50.592	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
Data File : VX037746.D
Acq On : 15 Sep 2023 10:13
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Quant Time: Sep 16 04:29:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:24:04 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
 Data File : VX037746.D
 Acq On : 15 Sep 2023 10:13
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Sep 16 04:29:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:24:04 2023
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 259344
Ion Ratio Lower Upper
168 100
99 55.1 56.6 85.0#



#2
Dichlorodifluoromethane
Concen: 49.117 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 85 Resp: 134265
Ion Ratio Lower Upper
85 100
87 32.7 16.7 50.1









#3
Chloromethane
Concen: 46.217 ug/l
RT: 1.300 min Scan# 35
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 137461
Ion Ratio Lower Upper
50 100
52 33.7 25.5 38.3



#4
Vinyl Chloride
Concen: 49.108 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 62 Resp: 157573
Ion Ratio Lower Upper
62 100
64 32.8 25.4 38.0









#5
Bromomethane
Concen: 50.406 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 115789
Ion Ratio Lower Upper
94 100
96 95.7 74.6 111.8



#6
Chloroethane
Concen: 47.992 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.007 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 64 Resp: 106087
Ion Ratio Lower Upper
64 100
66 32.1 25.1 37.7









#7
Trichlorofluoromethane
Concen: 50.343 ug/l
RT: 1.886 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 259698
Ion Ratio Lower Upper
101 100
103 65.1 51.4 77.2



#8
Diethyl Ether
Concen: 48.223 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 74 Resp: 96376
Ion Ratio Lower Upper
74 100
45 87.4 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.699 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 130109
Ion Ratio Lower Upper
101 100
85 44.4 37.0 55.6
151 76.0 58.6 87.8



#10
Methyl Iodide
Concen: 51.831 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion:142 Resp: 179570
Ion Ratio Lower Upper
142 100
127 45.8 41.6 62.4
141 14.5 11.6 17.4









#11
Tert butyl alcohol
Concen: 241.962 ug/l
RT: 2.965 min Scan# 308
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 176968
Ion Ratio Lower Upper
59 100
57 10.4 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 49.992 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 96 Resp: 126487
Ion Ratio Lower Upper
96 100
61 155.9 144.6 216.8
98 66.2 50.4 75.6









#13
Acrolein
Concen: 242.726 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 175380
Ion Ratio Lower Upper
56 100
55 70.3 57.1 85.7



#14
Allyl chloride
Concen: 49.314 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 41 Resp: 207806
Ion Ratio Lower Upper
41 100
39 69.2 55.8 83.6
76 37.8 24.3 36.5









#15
Acrylonitrile
Concen: 248.462 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 53 Resp: 404033
Ion Ratio Lower Upper
53 100
52 82.6 66.0 99.0
51 36.0 29.8 44.6



#16
Acetone
Concen: 228.873 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 43 Resp: 375160
Ion Ratio Lower Upper
43 100
58 32.9 21.9 32.9#









#17
Carbon Disulfide
Concen: 49.847 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 316209
Ion Ratio Lower Upper
76 100
78 9.3 7.0 10.6



#18
Methyl Acetate
Concen: 48.906 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 43 Resp: 289189
Ion Ratio Lower Upper
43 100
74 25.6 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 48.992 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 455989
Ion Ratio Lower Upper
73 100
57 21.5 17.3 25.9



#20
Methylene Chloride
Concen: 49.554 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 84 Resp: 150172
Ion Ratio Lower Upper
84 100
49 112.9 103.6 155.4
51 35.4 33.8 50.6
86 63.5 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 50.026 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 140549
Ion Ratio Lower Upper
96 100
61 134.2 119.2 178.8
98 64.4 48.7 73.1



#22
Diisopropyl ether
Concen: 49.805 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 45 Resp: 437172
Ion Ratio Lower Upper
45 100
43 71.1 69.7 104.5
87 31.7 20.8 31.2#
59 11.9 9.0 13.4









#23
 Vinyl Acetate
 Concen: 255.031 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX037746.D
 Acq: 15 Sep 2023 10:13




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 1646586
 Ion Ratio Lower Upper
 43 100
 86 11.6 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 48.958 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.000 min
 Lab File: VX037746.D
 Acq: 15 Sep 2023 10:13

Tgt Ion: 63 Resp: 257515
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.3 9.8
 100 4.5 2.1 6.5





#25

2-Butanone

Concen: 245.222 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.012 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 43 Resp: 567078

Ion Ratio Lower Upper

43 100

72 29.3 19.4 29.2#



#26

2,2-Dichloropropane

Concen: 49.207 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.000 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion: 77 Resp: 209137

Ion Ratio Lower Upper

77 100

97 24.1 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 49.776 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 96 Resp: 171445

Ion Ratio Lower Upper

96 100

61 136.2 0.0 312.2

98 63.6 0.0 129.0



#28

Bromochloromethane

Concen: 47.683 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion: 49 Resp: 117007


Ion Ratio Lower Upper

49 100

129 2.2 0.0 3.8

130 86.5 57.5 86.3#





#29

Tetrahydrofuran

Concen: 248.957 ug/l

RT: 5.007 min Scan# 643

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 358180

Ion Ratio Lower Upper

42 100

72 50.2 34.2 51.4

71 46.8 32.2 48.2



#30

Chloroform

Concen: 49.427 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.000 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion: 83 Resp: 280064

Ion Ratio Lower Upper

83 100

85 65.1 50.3 75.5









#31
Cyclohexane
Concen: 50.989 ug/l
RT: 5.470 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 222993
Ion Ratio Lower Upper
56 100
69 30.5 23.4 35.2
84 88.3 65.9 98.9



#32
1,1,1-Trichloroethane
Concen: 51.226 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 97 Resp: 245770
Ion Ratio Lower Upper
97 100
99 65.1 51.1 76.7
61 41.8 36.4 54.6









#33
1,2-Dichloroethane-d4
Concen: 49.897 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 65 Resp: 170017
Ion Ratio Lower Upper
65 100
67 53.6 0.0 100.2



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 114 Resp: 430033
Ion Ratio Lower Upper
114 100
63 19.1 0.0 45.0
88 15.5 0.0 36.6









#35
Dibromofluoromethane
Concen: 50.502 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:113 Resp: 145313
Ion Ratio Lower Upper
113 100
111 103.0 82.9 124.3
192 19.2 13.3 19.9



#36
1,1-Dichloropropene
Concen: 49.923 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 75 Resp: 197883
Ion Ratio Lower Upper
75 100
110 37.4 17.0 50.9
77 31.4 24.7 37.1





#37

Ethyl Acetate

Concen: 51.031 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.018 min

Lab File: VX037746.D


Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050






Tgt Ion: 43 Resp: 217161

Ion Ratio Lower Upper

43 100

61 14.6 10.6 16.0

70 11.3 8.3 12.5



#38

Carbon Tetrachloride

Concen: 51.159 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion: 117 Resp: 212170


Ion Ratio Lower Upper

117 100

119 94.8 78.2 117.2

121 31.4 24.4 36.6









#39
Methylcyclohexane
Concen: 50.346 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 253272
Ion Ratio Lower Upper
83 100
55 73.3 64.6 97.0
98 46.6 36.7 55.1



#40
Benzene
Concen: 49.662 ug/l
RT: 6.037 min Scan# 812
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 78 Resp: 605328
Ion Ratio Lower Upper
78 100
77 23.8 19.3 28.9









#41
Methacrylonitrile
Concen: 49.813 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.025 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 41 Resp: 110678
Ion Ratio Lower Upper
41 100
39 56.6 45.2 67.8
67 83.8 52.6 78.8#
52 34.1 22.7 34.1#



#42
1,2-Dichloroethane
Concen: 48.785 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.012 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 62 Resp: 215577
Ion Ratio Lower Upper
62 100
98 9.3 0.0 16.0









#43
Isopropyl Acetate
Concen: 50.707 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 353314
Ion Ratio Lower Upper
43 100
61 20.3 15.2 22.8
87 14.1 9.8 14.6



#44
Trichloroethene
Concen: 47.606 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion:130 Resp: 161598
Ion Ratio Lower Upper
130 100
95 99.7 0.0 204.4









#45
1,2-Dichloropropane
Concen: 49.103 ug/l
RT: 7.433 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 63 Resp: 155431
Ion Ratio Lower Upper
63 100
65 31.2 24.3 36.5



#46
Dibromomethane
Concen: 50.222 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 93 Resp: 114355
Ion Ratio Lower Upper
93 100
95 85.0 66.6 100.0
174 97.3 69.4 104.0





#47

Bromodichloromethane

Concen: 51.515 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 216033

Ion Ratio Lower Upper

83 100

85 64.4 49.8 74.6

127 9.2 7.1 10.7



#48

Methyl methacrylate

Concen: 50.902 ug/l

RT: 7.696 min Scan# 1084

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion: 41 Resp: 169547


Ion Ratio Lower Upper

41 100

69 92.5 59.2 88.8

39 61.7 47.7 71.5





#49

1,4-Dioxane

Concen: 966.542 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.012 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 84402

Ion Ratio Lower Upper

88 100

43 28.7 0.0 0.0#

58 65.6 60.1 90.1



#50

Toluene-d8

Concen: 51.121 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion: 98 Resp: 546543

Ion Ratio Lower Upper

98 100

100 64.6 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 252.149 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 1114690
Ion Ratio Lower Upper
43 100
58 41.5 30.4 45.6



#52
Toluene
Concen: 50.723 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 92 Resp: 398453
Ion Ratio Lower Upper
92 100
91 173.5 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 53.372 ug/l
 RT: 8.976 min Scan# 1294
 Delta R.T. -0.012 min
 Lab File: VX037746.D
 Acq: 15 Sep 2023 10:13




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 75 Resp: 232974
 Ion Ratio Lower Upper
 75 100
 77 31.6 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 51.867 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.006 min
 Lab File: VX037746.D
 Acq: 15 Sep 2023 10:13

Tgt Ion: 75 Resp: 251705
 Ion Ratio Lower Upper
 75 100
 77 31.4 25.3 37.9
 39 45.7 46.7 70.1#









#55
1,1,2-Trichloroethane
Concen: 50.526 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 97 Resp: 167342




Ion	Ratio	Lower	Upper
97	100		
83	86.8	65.8	98.8
85	56.2	41.6	62.4
99	63.8	47.8	71.8




#56
Ethyl methacrylate
Concen: 51.602 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 69 Resp: 255142

Ion	Ratio	Lower	Upper
69	100		
41	61.9	56.8	85.2
39	37.0	32.0	48.0





#57

1,3-Dichloropropane

Concen: 49.730 ug/l

RT: 9.311 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 266525

Ion Ratio Lower Upper

76 100

78 32.5 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 246.430 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion: 63 Resp: 568955

Ion Ratio Lower Upper

63 100

106 28.7 22.5 33.7





#59

2-Hexanone

Concen: 251.303 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 43 Resp: 864988

Ion Ratio Lower Upper

43 100

58 56.5 26.2 78.5



#60

Dibromochloromethane

Concen: 52.971 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.006 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion: 129 Resp: 174770

Ion Ratio Lower Upper

129 100

127 77.9 38.8 116.4









#61
1,2-Dibromoethane
Concen: 51.226 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 179632
Ion Ratio Lower Upper
107 100
109 95.0 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 50.619 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 95 Resp: 196656
Ion Ratio Lower Upper
95 100
174 82.2 0.0 135.6
176 79.4 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:117 Resp: 410931
Ion Ratio Lower Upper
117 100
82 53.7 46.2 69.4
119 31.5 25.5 38.3



#64
Tetrachloroethene
Concen: 48.450 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion:164 Resp: 135780
Ion Ratio Lower Upper
164 100
166 125.3 102.7 154.1
129 94.2 80.4 120.6
131 92.4 76.5 114.7









#65
Chlorobenzene
Concen: 48.504 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:112 Resp: 429927
Ion Ratio Lower Upper
112 100
114 32.2 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 51.725 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion:131 Resp: 162476
Ion Ratio Lower Upper
131 100
133 93.5 47.1 141.4
119 65.4 32.2 96.6









#67
Ethyl Benzene
Concen: 50.061 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 769315
Ion Ratio Lower Upper
91 100
106 31.7 24.9 37.3



#68
m/p-Xylenes
Concen: 101.121 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 106 Resp: 600485
Ion Ratio Lower Upper
106 100
91 202.4 165.5 248.3





#69

o-Xylene

Concen: 50.233 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:106 Resp: 299129

Ion Ratio Lower Upper

106 100

91 213.0 108.9 326.8



#70

Styrene

Concen: 51.686 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion:104 Resp: 494103


Ion Ratio Lower Upper

104 100

78 50.2 43.6 65.4

103 55.9 43.6 65.4





#71

Bromoform

Concen: 52.834 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:173 Resp: 130939

Ion Ratio Lower Upper

173 100

175 48.3 24.1 72.4

254 0.1 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion:152 Resp: 205192


Ion Ratio Lower Upper

152 100

115 85.1 43.3 129.9

150 172.5 0.0 345.0





#73

Isopropylbenzene

Concen: 51.577 ug/l

RT: 10.963 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:105 Resp: 776441

Ion Ratio Lower Upper

105 100

120 26.7 12.9 38.6



#74

N-ethyl acetate

Concen: 52.428 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Tgt Ion: 43 Resp: 319637




Ion Ratio Lower Upper


43 100

70 43.6 32.9 49.3

55 24.7 19.3 28.9

61 24.9 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 49.495 ug/l

RT: 11.213 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 262045

Ion Ratio Lower Upper

83 100

131 10.2 5.3 15.8

85 65.0 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 57.246 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion: 75 Resp: 287462

Ion Ratio Lower Upper

75 100

77 32.1 19.1 57.5





#77

Bromobenzene

Concen: 50.455 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:156 Resp: 187998

Ion Ratio Lower Upper

156 100

77 154.3 88.4 265.3

158 98.6 48.0 144.0



#78

n-propylbenzene

Concen: 52.838 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion: 91 Resp: 898303

Ion Ratio Lower Upper

91 100

120 23.2 11.5 34.4









#79
2-Chlorotoluene
Concen: 50.690 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 531369
Ion Ratio Lower Upper
91 100
126 34.8 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 52.237 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 105 Resp: 664087
Ion Ratio Lower Upper
105 100
120 47.8 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 53.480 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 79232
Ion Ratio Lower Upper
75 100
53 96.6 90.5 135.7
89 46.9 36.9 55.3



#82
4-Chlorotoluene
Concen: 51.534 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 91 Resp: 608183
Ion Ratio Lower Upper
91 100
126 30.4 14.2 42.6





#83

tert-Butylbenzene

Concen: 52.282 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 671758

Ion Ratio Lower Upper

119 100

91 57.0 31.2 93.6

134 22.3 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 52.022 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.006 min

Lab File: VX037746.D




Acq: 15 Sep 2023 10:13


Tgt Ion:105 Resp: 658015

Ion Ratio Lower Upper

105 100

120 43.9 21.2 63.6





#85

sec-Butylbenzene

Concen: 52.752 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:105 Resp: 822306

Ion Ratio Lower Upper

105 100

134 19.9 9.8 29.4



#86

p-Isopropyltoluene

Concen: 52.579 ug/l

RT: 12.012 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion:119 Resp: 681771


Ion Ratio Lower Upper

119 100

134 26.5 12.8 38.3

91 24.2 13.1 39.3









#87
1,3-Dichlorobenzene
Concen: 49.801 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 345898
Ion Ratio Lower Upper
146 100
111 40.5 22.1 66.1
148 63.8 31.8 95.3



#88
1,4-Dichlorobenzene
Concen: 48.586 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion:146 Resp: 344312
Ion Ratio Lower Upper
146 100
111 39.5 21.3 63.7
148 64.9 31.8 95.4









#89
n-Butylbenzene
Concen: 53.164 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 604859
Ion Ratio Lower Upper
91 100
92 52.3 27.4 82.0
134 25.9 12.6 37.8



#90
Hexachloroethane
Concen: 53.292 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion: 117 Resp: 117268
Ion Ratio Lower Upper
117 100
201 75.6 32.9 98.6





#91

1,2-Dichlorobenzene

Concen: 48.832 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:146 Resp: 346286

Ion Ratio Lower Upper

146 100

111 41.4 22.4 67.0

148 63.6 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 52.579 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. 0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion: 75 Resp: 55489


Ion Ratio Lower Upper

75 100

155 89.1 36.1 108.5

157 113.3 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 51.125 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.006 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:180 Resp: 213559

Ion Ratio Lower Upper

180 100

182 96.0 48.3 144.8

145 31.7 18.0 54.0



#94

Hexachlorobutadiene

Concen: 51.017 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX037746.D

Acq: 15 Sep 2023 10:13




Tgt Ion:225 Resp: 99452


Ion Ratio Lower Upper

225 100

223 61.9 31.1 93.2

227 61.4 32.5 97.5









#95
Naphthalene
Concen: 52.266 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:128 Resp: 738377
Ion Ratio Lower Upper
128 100
127 13.0 10.3 15.5
129 11.1 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 50.592 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX037746.D
Acq: 15 Sep 2023 10:13

Tgt Ion:180 Resp: 217481
Ion Ratio Lower Upper
180 100
182 95.4 46.9 140.6
145 33.0 17.8 53.5

